



Certificate of Compliance

Certificate: 2268167 (LR 44340C)

Master Contract: 162029

Project: 2559152

Date Issued: October 24, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, AScT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Hand-held digital multimeter, Models 15XP-A, 15XP-B and 15XP-x, internally powered 1x9V NEDA, rated CAT III 600 V.

Notes:

1. The above models are Equipment Class II, Pollution Degree 2, Measurement CATIII 600V.
2. The suffix "x" can be 0-9 or A-Z.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements Requirements + Update No. 1:2008



Certificate: 2268167 (LR 44340C)

Master Contract: 162029

Project: 2559152

Date Issued: October 24, 2012

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements