

# Certificate of Compliance

**Certificate:** 1446924

**Master Contract:** 162029 (LR 44340C)

**Edition:** 1


**Date Issued:** October 30, 2003

**Issued to:** Fluke Corp  
PO Box 9090  
Everett, WA 98206-9090  
USA

*The products listed below are eligible to bear the CSA Mark shown  
with adjacent indicators 'C' and 'US'*



**Issued by:**   
Stephane Poutissou, P. Eng.

**Approved by:**   
Terry Nagy  
Operations Manager

## PRODUCTS

CLASS 3631 04 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 84 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE (Certified to U.S. Standards)

Hand-held Digital Multimeters, Models 5XP, 750 V ac or 1000 V dc, Installation Category I (1000 V), Installation Category II (600 V) and Installation Category III (300 V); Pollution Degree 2, current measuring circuit 0.01 mA to 200 mA max.

Hand-held Digital Multimeters, Models 15XP, 750 V ac or 1000 V dc, Installation Category I (1000 V), Installation Category II (600 V) and Installation Category III (300 V); Pollution Degree 2, current measuring circuit 0.01 mA to 2 mA max.

Hand-held Digital Multimeters, Models 35XP, 750 V ac or 1000 V dc, Installation Category I (1000 V), Installation Category II (600 V) and Installation Category III (300 V); Pollution Degree 2, current measuring circuit 0.01 mA to 2 mA max.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.



CSA INTERNATIONAL

**Certificate:** 1446924

**Master Contract:** 162029 (LR 44340C)

**Edition:** 1

**Date:** October 30, 2003

**APPLICABLE REQUIREMENTS**

- CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)
- CAN/CSA-C22.2 No. 1010.1B-97 - Amendment 2 to CAN/CSA-C22.2 No. 1010.1-92, "Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements"
- UL Std No. 61010B-1 - Equipment for Measurement Use; Part I: General Requirements
- ISA S82.01-1994 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling, and Related Equipment - General Requirements

**CONDITIONS OF ACCEPTABILITY**

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.



## *Supplement to Certificate of Compliance*

**Certificate:** 1446924

**Master Contract:** 162029 (LR 44340C)

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Edition</b>	<b>Date</b>	<b>Description</b>
1	October 30, 2003	Original Certification.