



# Certificate of Compliance

**Certificate:** 1446924 (LR 44340C)

**Master Contract:** 162029

**Project:** 2559153

**Date Issued:** November 27, 2012

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

Attention: W. John Hyacinthe

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Paul Tanaka*

**Issued by:** Paul Tanaka, ASCT.

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Hand-held Digital Multimeters, Model 5XP, internally powered 1x9V NEDA; rated CATII 600V, CATIII 300 V; current measuring circuit 0.01 mA to 200 mA max.

Hand-held Digital Multimeters, Models 35XP and 35XPA, internally powered 1x9V NEDA; rated CATII 600V, CATIII 300V; current measuring circuit 0.01 mA to 2 mA max.

## **Notes:**

- The above models are Equipment Class II, Pollution Degree 2.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + Update No. 1:2008



**Certificate:** 1446924 (LR 44340C)

**Master Contract:** 162029

**Project:** 2559153

**Date Issued:** November 27, 2012

---

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

#### Reference Standards

IEC 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements