



Certificate of Compliance

Certificate: 1549885 (LR 44340)

Master Contract: 162029

Project: 1885856

Date Issued: 2007/02/26

Issued to: **Fluke Corporation**
P.O. Box 9090
Everett, WA 98206-9090
USA
Attention: Mr. Dan Cordwell

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Mehrdad Sadeghieh, P. Eng.

Authorized by: Ken Rutledge, Product Group Manager

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT
CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Clamp meters, Models AC71, AC71A, AC71B and AC71C, battery operated, rated 600 A, duty cycle 10 min on, 10 min off, 750 V ac, 1000V dc, CAT III 600 V, CAT II 1000 V.

Note: The above models are Equipment Class 2, Pollution Degree 2.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1549885 (LR 44340)

Master Contract: 162029

Project: 1885856

Date Issued: 2007/02/26

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010.1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-032-04 - Safety requirements for electrical equipment for measurement, control, and laboratory use. Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement; IEC No. 61010-2-032: 2002