



# Certificate of Compliance

**Certificate:** 1454482 (LR 44340C)

**Master Contract:** 162029

**Project:** 2557010

**Date Issued:** October 22, 2012

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

Attention: W. John Hyacinthe

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Paul Tanaka*

**Issued by:** Paul Tanaka, ASCT.

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Clamp meter, Model AC75, AC75A, AC75B and AC75C, battery operated, rated AC Clamp, 500 A continuous, 500-600 A duty cycle 10 min on, 10 min off. DC A 400  $\mu$ A, AC V 750 V, DC V 1000 V, 600 V CAT III, 1000 V CAT II.

## **Notes:**

The above model is Equipment Class 2, Pollution Degree 2, Measurement CAT III 600V / CATII 1000V

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + Update No. 1:2008



**Certificate:** 1454482 (LR 44340C)

**Master Contract:** 162029

**Project:** 2557010

**Date Issued:** October 22, 2012

---

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-032-04 - Safety requirements for electrical equipment for measurement, control, and laboratory use. Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement; IEC No. 61010-2-032: 2002

#### REFERENCE STANDARDS

Compliance to the following Standards and Bulletin was also assessed:

IEC 61010-1, Edition 2 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-032, Edition 2 - Safety requirements for electrical equipment for measurement, control, and laboratory use. Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement.

CB Bulletin OC, No 110a - Adherence to IEC Standards - National differences for IEC 61010-1, Edition 2 and IEC 61010-2-032, Edition 2.