



Certificate of Compliance

Certificate: 2226784 (LR 44340C)

Master Contract: 162029

Project: 2559151

Date Issued: October 24, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, ASCT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Clamp Meter, Models ACD-20SW, ACD-21SW, ACD-21SWC, ACD-22SW and ACD-23SW; internally powered by 2 x 1.5 AAA batteries, Type A, CAT III 600 V.

Notes:

The above models are Equipment Class II, Pollution Degree 2, Measurement Category III.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD)

CAN/CSA-C22.2 No. 61010-2 -032:04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (CEI/IEC 61010-2-032:2002)

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 2226784 (LR 44340C)

Master Contract: 162029

Project: 2559151

Date Issued: October 24, 2012

Reference Standards

IEC 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + Update No. 1:2008

IEC 61010-2-032:2002 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement