



Certificate of Compliance

Certificate: 2226732

Master Contract: 162029

Project: 2566117

Date Issued: October 23, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, ASCT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Industrial TRMS Clamp Meter, Models ACD-3300 IND and ACDC-3400 IND; internally powered by 1 x 9Vdc NEDA 1604 IEC 6F22 battery, Type A, CAT III 1000 V, CAT IV 600 V, 1000 A ac/dc.

Notes:

The above models are Equipment Class II, Pollution Degree 2, Measurement CATIII 1000V, CATIV 600V.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD)+ Update No. 1:2008

CAN/CSA-C22.2 No. 61010-2 -032:04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (CEI/IEC 61010-2-032:2002)

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 2226732

Master Contract: 162029

Project: 2566117

Date Issued: October 23, 2012

Reference Standards

IEC 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

IEC 61010-2-032:2002 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 2: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement