

VERIFICATION

of conformity with Low Voltage Directive

Verification No.: ACT202263LVD-A1

Document holder:
Amprobe Test Tools

Address:
6920 Seaway Blvd., Everett WA 98203, USA

Trade mark:


Type of product:
Digital Clamp Meter

Type designation:
AMPROBE ACDC-54NAV,
AMPROBE ACD-53NAV,
AMPROBE ACDC-52NAV,
AMPROBE ACD-51NAV,
AMPROBE ACD-50NAV

Technical data:
Type A, Class II
Powered by a 9V battery (Type 1604)
CAT III / 1000V and CAT IV / 600V
1000Aac for AMPROBE ACD-53NAV,
1000Aac/dc for AMPROBE ACDC-54NAV,
600Aac for AMPROBE ACD-51NAV and
AMPROBE ACD-50NAV
600Aac/dc for AMPROBE ACDC-52NAV

A sample of the product has been assessed with respect to CE-marking according to the Low Voltage Directive (2006/95/EC) and found to comply with the essential requirements of the Directive.

The Standard(s) used for showing the compliance and the full details of the results are given in the Test Report as detailed below:

Standard(s)	Report No.	Report Issued Date
IEC / EN 61010-2-032:2002 (2nd Edition) and IEC / EN 61010-1:2001 (2nd Edition)	ACT202263-A1	March 30, 2011

The holder of the verification is authorized to use this verification in connection with the EC declaration of conformity according to the Directive. The CE marking may only be used if all relevant and effective EC Directives are complied with. Together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the Low Voltage Directive.

Approved by:



Vincent Tan
Acts Certification and Testing Services
March 30, 2011



Acts Certification and Testing Services Co., Ltd.

1st Floor, No. 20, Lane 61, Tianshiang Road, Taipei, Taiwan Tel: +886-2-25858775 Fax: +886-2-25983002