



CSA INTERNATIONAL

# Certificate of Compliance

**Certificate:** 1671838 (LR 44340C)

**Master Contract:** 162029

**Project:** 1885851

**Date Issued:** 2007/02/26

**Issued to:** Fluke Corporation

P.O. Box 9090

Everett, WA 98206-9090

USA

Attention: Mr. Dan Cordwell

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'*



**Issued by:** Mehrdad Sadeghieh, P. Eng.

**Authorized by:** Ken Rutledge, Product Group Manager

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Clamp meters, battery operated, CAT III, Models AC40A and AC40B, rated AC Clamp, 400 A, 600 V ac and Input Terminals 600 V ac/dc; Models AD40A and AD40B, rated AC Clamp, 400 A, 600 V ac.

Note: The above model is Equipment Class 2, Pollution Degree 2.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



**Certificate:** 1671838 (LR 44340C)

**Master Contract:** 162029

**Project:** 1885851

**Date Issued:** 2007/02/26

---

### **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010.1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-32-04 - Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

IEC 61010-2-032(2002) - Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement