



Certificate of Compliance

Certificate: 2565106

Master Contract: 162029

Project: 2565106

Date Issued: January 21, 2013

Issued to: Fluke Corporation
6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Dean Chen

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Lotus Yao

Issued by: Lotus Yao

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Hand-held Digital Multimeter, Model AM-420, with integral test leads, rated CAT II 250V. Battery powered by a 6F22 (carbon - zinc) battery.

Notes:

The above model is Equipment Class II, Pollution Degree 2.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements + CAN/CSA C22.2 No 61010-1-04 Update No. 1:2008



Certificate: 2565106

Master Contract: 162029

Project: 2565106

Date Issued: January 21, 2013

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-031-07 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 31: Safety requirements for hand-held probe assemblies for electrical measurement and test

UL 61010-031 (1st Edition - 2007) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test



Supplement to Certificate of Compliance

Certificate: 2565106

Master Contract: 162029

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
2565106	Jan 21, 2013	Original Certification.