



# Certificate of Compliance

**Certificate:** 2571588

**Master Contract:** 162029

**Project:** 2571588

**Date Issued:** April 9, 2013

**Issued to:** Fluke Corporation

6920 Seaway Blvd

Everett, WA 98203

USA

Attention: Dean Chen

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Jerry Ning*

**Issued by:** Jerry Ning

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Hand-held Digital Multimeter, Model AM-450, rated 9V, to be powered by a 6F22 (carbon-zinc) battery.

### Notes:

- The above model is Equipment Class II, Pollution Degree 2, measurement CAT III 600V.
- Provided with Accessory Test Lead, Model TL175.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements +  
CAN/CSA C22.2 No 61010-1-04 Update No. 1:2008



**Certificate:** 2571588

**Master Contract:** 162029

**Project:** 2571588

**Date Issued:** April 9, 2013

---

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement,  
Control, and Laboratory Use - Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 2571588

**Master Contract:** 162029

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
2571588	Apr 9, 2013	Original Certification