



Certificate of Compliance

Certificate: 2455689

Master Contract: 162029

Project: 2587729

Date Issued: December 18, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, AScT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Hand-held digital multimeter: 8.9 A continuous operation and 10 A at 20 min ON/10 min OFF Cycling, battery operated:

- AM-570 rated CAT III 1000 V; CAT IV 600 V
- AM-560 rated CAT III 1000 V; CAT IV 600 V
- AM-530 rated CAT III 600 V
- AM-520 rated CAT III 600 V
- AM-510 rated CAT III 600 V

Notes:

The above models are Equipment Class II, Pollution Degree 2.



Certificate: 2455689

Master Contract: 162029

Project: 2587729

Date Issued: December 18, 2012

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD) + Update No. 1:2008

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

ISA-82.02.01 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment – General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements