



# Certificate of Compliance

**Certificate:** 2491030

**Master Contract:** 162029

**Project:** 2491030

**Date Issued:** July 20, 2012

**Issued to:** Fluke Corporation

6920 Seaway Blvd  
Everett, WA 98203  
USA

Attention: W. John Hyacinthe

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



*Cindy Johnson*

**Issued by:** Cindy Johnson, ASCT.

## **PRODUCTS**

**CLASS 3631 05** - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

**CLASS 3631 85** - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Handheld digital multi-meter, Models AM-500 and AM-500-EUR, rated: 2 x 1.5V Alkaline AAA batteries, CAT III, 600 Vac/dc, 8A continuous / 10 A 1 min ON/10 min OFF.

Note: The above models are Equipment Class II, Pollution Degree 2.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD) + CSA Update No. 1:2008

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



**Certificate:** 2491030

**Master Contract:** 162029

**Project:** 2491030

**Date Issued:** July 20, 2012

---

ISA-82.02.01 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment – General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements