



Certificate of Compliance

Certificate: 70017850

Master Contract: 162029

Project: 70017850

Date Issued: March 19, 2015

Issued to: **Fluke Corporation**
6920 Seaway Blvd
Everett, WA 98203
USA
Attention: Dean Chen

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Lotus Yao

Issued by: Lotus Yao

PRODUCTS

CLASS 3631 06 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 86 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Clamp Meter, Models AMP-25 and AMP-25-EUR, battery operated, 2 x 1.5 Vdc (LR44).

Notes:

1. The above models are Equipment Class II, Pollution Degree 2, Measurement Category CAT III 600 V, 300 A dc/ac.
2. Mode of operation: Continuous.
3. Environmental Conditions: Extended: 0 to 50 °C, 2000 m max, up to 80% R.H. when temperatures up to 30 °C, up to 75% R.H. when temperatures above 30 to 40 °C, up to 45% R.H. when temperatures above 40 to 50 °C as specified by manufacturer.



Certificate: 70017850

Master Contract: 162029

Project: 70017850

Date Issued: March 19, 2015

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-030-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

UL Standards:

UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

UL Std. No. 61010-2-030 (1st Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-030: Particular requirements for testing and measuring circuits

Reference Standards:

IEC 61010-2-032:2012 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 2-032: Particular requirements for hand-held and hand-manipulated current sensors for electrical test and measurement



Supplement to Certificate of Compliance

Certificate: 70017850

Master Contract: 162029

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
70017850	Mar 19, 2015	Original Certification.