



# EC Declaration of Conformity

Document number

CE329

We Fluke Precision Measurement Ltd, 52 Hurricane Way, Airport Industrial Estate, Norwich, Norfolk, NR6 6JB UK.  
Certify that the product/s described below is/are in conformity with the following directives:

**73/23/EEC "Low Voltage Directive for Electrical Safety amended by the directive 93/68/EEC"**

**89/336/EEC "Electromagnetic Compatibility Directive amended by the directives 92/31/EEC, 91/263/EEC, 93/68/EEC, 93/97/EEC, 93/68/EEC, 96/C285/06 and 91/13/EC"**

Model	Order Number	Description
ISO Plus	SI1310Z	Multimeter & Insulation Tester
ISO Plus M	SI1311Z	Multimeter & Insulation Tester

The product/s has/have been assessed by application of the following standards or specifications:

### Low Voltage Directive

**BS EN 61010-1 : 2001**

Safety requirements for electrical equipment for measurement, control and laboratory use.

600V  
Cat III  
Pollution Degree 2

**BS EN 61010-031 : 2002**

Safety requirements for hand-held probe assemblies for electrical measurement and test.

### Electromagnetic Compatibility Directive

**BS EN 61326 : 1998 +A1, A2 & A3**

Electrical Equipment for measurement, control and laboratory use – EMC requirements.

Annex C  
Performance Criterion B

I certify for and on behalf of Fluke Precision Measurement Ltd. that the product/s mentioned above conform/s to the requirements mentioned above.

Signature		Name	Desmond Ebenezer
Position	Technology Manager	Date	22 <sup>nd</sup> January 2007