



Certificate of Compliance

Certificate: 1743022 (LR 44340C)

Master Contract: 162029

Project: 2557014

Date Issued: October 23, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, AScT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeter, Handheld, battery operated, Models DM7B & DM7C, 9 V dc, measurement range 600 V ac/dc, 200 mA dc and Models DM9B & DM9C, 2 x 1.5 V dc, measurement range 600 V ac/dc, 5 A ac/dc continuous and 10 A ac/dc intermittent 2 min ON, 20 min OFF.

Notes:

The above model is Pollution Degree 2, Measurement Category II.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements+ Update No. 1:2008

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 1743022 (LR 44340C)

Master Contract: 162029

Project: 2557014

Date Issued: October 23, 2012

Reference Standards:

Compliance to the following Standard and Bulletin was also assessed:

IEC 61010-1, Edition 2 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements