



Certificate of Compliance

Certificate: 1872245 (LR44340C)

Master Contract: 162029

Project: 2566118

Date Issued: December 18, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Paul Tanaka

Issued by: Paul Tanaka, AScT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Digital Multimeters, Handheld, Amprobe Models HD110C and HD160C, battery operated, 9 V dc, measurement range 1000 V ac, 1500 Vdc, 2 A ac/dc

Notes:

The above models are Equipment Class II, Pollution Degree 2, Measurement Category IV 1000V.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements+Update No. 1:2008*

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 1872245 (LR44340C)

Master Contract: 162029

Project: 2566118

Date Issued: December 18, 2012

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

* Update No. 1:2008 applies only to Model HD110C