



Certificate of Compliance

Certificate: 2470549

Master Contract: 162029

Project: 2470549

Date Issued: March 5, 2012

Issued to: Fluke Corporation

6920 Seaway Blvd
Everett, WA 98203
USA

Attention: W. John Hyacinthe

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Cindy Johnson

Issued by: Cindy Johnson, ASCT.

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Wiring Inspection Tester, Model INSP-3, hand-held, battery operated 1 x 9V.

Note: The above model is Equipment Class II, Pollution Degree 2, Measurement CAT III, 95-120 Vac, 50 or 60 Hz.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (IEC 61010-1:2001, MOD)

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Certificate: 2470549

Master Contract: 162029

Project: 2470549

Date Issued: March 5, 2012

ISA-82.02.01 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment – General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements