

Certificate of Compliance

Certificate: 2254096

Master Contract: 248931

Project: 2254096

Date Issued: December 29, 2009


Issued to: **Fluke Corporation**
6920 Seaway Blvd Everett,
WA 98203
USA

Attention: John Wesley Hyacinthe

The products listed below are eligible to bear the CSA Mark shown



Issued by: Lotus Yao



Authorized by: Anthony Lee, P. Eng
Certification Specialist - Asia

PRODUCTS

CLASS -3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS -3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE

Electrical Tester, double-insulated, Model PY-1A, rated 120/240/277/480 Vac, 0-65 Hz, and 160/330/380/600Vdc, 3.5 mA maximum.

Note: The above model is Class II Measurement Category III 600V equipment for indoor use in Pollution Degree 2, 5 to 40 deg C ambient, and up to 2000 m environment.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Std. No. 61010-1 (2nd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Certificate: 2254096

Project: 2254096



Master Contract: 248931

Date Issued: December 29, 2009

-
- CAN/CSA-C22.2 No. 61010-031-07 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 31: Safety requirements for hand-held probe assemblies for electrical measurement and test
 - UL 61010-031 (1st Edition - 2007) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test
 - IEC 61010-031 (Edition 1.1 2008-06) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 031: Safety requirements for hand-held probe assemblies for electrical measurement and test (used as a guide for the flexing/pull test only)



Supplement to Certificate of Compliance

Certificate: 2254096

Master Contract: 248931

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
2254096	2009-12-29	Multiple Listing for Fluke Corporation; Electrical Tester, double-insulated, Model PY-1A, rated 120/240/277/480 Vac, 0-65 Hz, and 160/330/380/600 Vdc, 3.5 mA maximum; Submitter Model 3020, Listee Model PY-1A.

CORRELATION OF MODEL DESIGNATIONS

SUBMITTOR'S MODEL	LISTEE'S MODEL
3020	PY-1A