

CERTIFICATE OF CONFORMITY

LVD Directive 2006/95/EC

Certificate No. 151079-01
Page 1/2
Date of Issue 2010-06-21

Certificate Holder FLUKE CORP
FLUKE ELECTRONICS PARK 6920 SEAWAY BLVD PO
BOX 9090 EVERETT WA 98206 USA

Manufacturer TENMARS ELECTRONICS CO LTD
6TH FL 586 RUI GUANG RD
HEIHU TAIPEI 114 TAIWAN

Production site TENMARS ELECTRONICS CO LTD
6F FL 2 258 LAND RUI GUANG RD
NEIHU TAIPEI 114 TAIWAN

Certified Product Analog Clamp Meter
Model RS-3 PRO, RS-1007 PRO

Trademark 

Rated Voltage / Frequency See appendix

Rated Current / Power See appendix

Insulation Class II

Degree of protection (IP) -

Tested acc. to EN 61010-1:2001 and EN 61010-2-032:2002

Test Report No. E214542-A2-IT-1 issue date 2010-06-08 with E214542-A2-IT-2 issue date 2010-06-21

Additional -

Certification Manager
Jan-Erik Storgaard

Representative models of the above noted product(s) have been submitted to Underwriters Laboratories Inc. and been found to be in compliance with the above mentioned harmonized standard(s) per Article 5 of *Directive 2006/95/EC of the European Parliament of the Council of 12 December 2006/95/EC on the harmonisation of the laws of Member States relating to electrical equipment designed for use within certain voltage limits.*

Per Article 5 of Directive 2006/95/EC, "electrical equipment which complies with the safety provisions of harmonized standards shall be regarded by their competent administrative authorities as complying with the provisions of Article 2, for the purposes of placing on the market and free movement as referred to in Articles 2 and 3 respectively."

As stated in Article 8 of Directive 2006/95/EC, "Before being placed on the market, the electrical equipment must have affixed to it the CE marking provided for in Article 10 attesting to its conformity to the provisions of this Directive, including the conformity assessment procedure described in Annex IV." This remains the responsibility of the manufacturer or his authorized representative established within the Community and is not covered as part of this certificate.

Certification Body UL International Demko A/S, Lyskaer 8, P.O. Box 514, DK-2730
Herlev, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com

Appendix CERTIFICATE OF CONFORMITY

Certificate No. 151079-01
Page 2/2
Date of Issue 2010-06-21

Rating:

Model: RS-3 PRO -
Battery Power: 1x1.5 V, AAA size.
Measurement Category: CAT IV 600 Va.c., 300 A.

Model: RS-1007 PRO -
Battery Power: 1x1.5 V, AAA size.
Measurement Category: CAT IV 600 Va.c., 1000A.

Certification Body

UL International Demko A/S, Lyskaer 8, P.O. Box 514, DK-2730
Herlev, Denmark, Tel. +45 44 85 65 65, info.dk@dk.ul.com
www.ul-europe.com



the standard in safety

Underwriters
Laboratories