



Certificate of Compliance

Certificate: 2528440

Master Contract: 162029

Project: 2528440

Date Issued: June 29, 2012

Issued to: Fluke Corp.
6920 Seaway Blvd.
Everett, WA 98203
USA

Attention: Mr. Joe Swanzy

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only



Issued by: 
Alvin Kuang, Certifier

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE (Certified to U.S. Standards)

Non-contact Voltage Probe, Models VP-1000B and VP-1000SB, rated 3 V dc (2 x AAA).

Note: The above models are Equipment Class II, Pollution Degree 2, Measurement Category IV (1000V).

APPLICABLE REQUIREMENTS

- | | | |
|---|---|--|
| CAN/CSA-C22.2 No. 61010-1-04 | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| UL Std. No. 61010-1 (2 nd Edition) | - | Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |



Supplement to Certificate of Compliance

Certificate: 2528440

Master Contract: 162029

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
2528440	June 29, 2012	Original Certification.