



# Certificate of Compliance

**Certificate:** 2528440

**Master Contract:** 162029

**Project:** 70044779

**Date Issued:** September 2, 2015

**Issued to:** Fluke Corp.  
6920 Seaway Blvd.  
Everett, WA 98203  
USA

**Attention:** Mr. Joe Swanzy

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only*



**Issued by:**

*Lotus Yao*  
Lotus Yao

## PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT

CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE (Certified to U.S. Standards)

Non-contact Voltage Probe, Models VP-1000, VP-1000B and VP-1000SB, rated 3 V dc (2 x AAA).

Note: The above models are Equipment Class II, Pollution Degree 2, Measurement Category IV(1000V).

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 + - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements  
Update No. 1: 2008

UL Std. No. 61010-1 (2<sup>nd</sup> Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 2528440

**Master Contract:** 162029

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70044779	September 2, 2015	Addition of Model VP-1000 and re-evaluations to CAN/CSA-C22.2 No. 61010-1-04 + Update No. 1: 2008.
2528440	June 29, 2012	Original Certification.